



Kingdom of Saudi Arabia
Ministry of Education
University of Hafr ALBatin



بسم الله الرحمن الرحيم
جامعة حفر الباطن
University of Hafr Al Batin

المملكة العربية السعودية
وزارة التعليم
جامعة حفر الباطن

نموذج 1 / مثال على طلب شراء (جدول مواصفات احتياج المعامل من المواد والأجهزة)

الكلية / قسم		اسم المعمل:		رقمه:	
م	المواصفات المطلوبة	الوحدة	الكمية	السعر الإجمالي	
				رقماً	كتابة
1.	PH meater Sensitivity 0.001pH unit Accuracy 0.001 pH unit Accept standard glass pH combination Auto	حبة	1		
2.	UV-VISDouble beam Spectrophotometer المواصفات: -Double beam. -Wavelength range: 190 to 1100 nm. - High resolution, Band pass: 1.5 nm - Scan speed: 10, 100, 200, 400, 800, 1200, 2400 and 3600 nm/min. - Detector: Silicon photodiode - (10)Cuvettes UV Quartz - Easy of use, High Precision Accuracy - Power requirements: 220 - 240 V AC, 50/60 Hz, 300 VA. Can be used in many applications from routine to Fine research. The U-2900 is equipped with a large 10.4" Color LCD Display and best stability and noise space in its class. Innovative memory compatibility with large data storage capabilities	جهاز	1		
3.	Conductivity/TDS meter المواصفات: -Bench Type. -Range Up to 200 mS/cm. -Auto and manual calibrated Auto rang capability with 4 standard -KCl solution -Memory up to 100 data sets -Accuracy 0.5% full scale+1digit. -Sensitivity 0.5 µS/cm. -With electrode holder -With KCL standard solution of 84 µS/cm, 1413 µS/cm, 12088 MS/CM, 11108mS/cm.	جهاز	1		